


CYCLOTRON OVERVIEW


- UWMCF houses the only fast neutron cyclotron particle accelerator being used for cancer treatment and research.
- A combination of a magnetic field and an oscillating electric field accelerate the particle in a circle.
- The two oscillating electric fields must be either completely in phase with each other or completely out of phase for the particles to accelerate.
- The RF phase meter is a critical component of the control system that measures the phase difference and reports the error.

PROBLEM STATEMENT

- RF Phase Meter a critical component in proton beam extraction and efficiency.
- Current RF phase meter was built in the 1980s. It has outdated components with decreasing reliability and accuracy.
- Goal: upgrade the RF Phase Meter with smaller and improved components and provide thorough documentation for the UWMCF staff.



REQUIREMENTS



- Output a DC voltage with slope 50mV/° centered at 0V.
- Operate at frequencies around 29.5MHz.
- Ability to switch between 0 degree and 180 degree operation modes.
- Accuracy within ± 5 mV ($\pm 0.1^\circ$) with long term stability.
- System must be self contained within a standard 19" rack.
- Ability to interface with the front pushbuttons and monitoring outputs.
- Provide thorough and accurate documentation for the UWMCF staff.



DESIGN PROCESS

- Designed and simulated circuits in LTSpice, then prototypes were built and tested on benchtop with function generators.
- RPM-50 prototype was tested with cyclotron system to ensure accurate phase detection.
- Two speciality PCBs, RPM-50 and RPM-52 were designed to consolidate circuitry and allow for future upgrades.
- A 19" rack was built to house power supplies, wiring and allow for interfacing. The two speciality PCBs were integrated into the rack.
- The RF Phase Meter was tested with the whole cyclotron system. It provided high proton beam extraction power and efficiency.

PCB AND RACK DESIGN

RPM-50

- RPM-50 is composed of four layers and provides signal conditioning to the output of the AD8302 IC. It provides the main system outputs (Phase Set and Phase Error).
- Features: inverting op amp and low noise relay to allow for mode switching, differential amplifier circuit for signal comparison, two trimmers for gain tuning.
- Stitching was incorporated to shield critical signals from EMI.




RPM-52

- The RPM-52 is composed of two layers and provides control flow for status interlocking to the cyclotron system.
- Features: multiple relays for signal control, internal/ external power separation through use of double-pole double-throw relays, manual control of operation modes through switching.

RACK DESIGN


- The RF Phase Meter is completely housed in a 19" rack.
- Features: AC Relay, three DC power supplies, terminal blocks, front and back panel connectors, interfacing for manual control, front panel door for box access.



CONCLUSION AND FUTURE WORK



- A system to detect the phase difference between two RF signals was built, tested and validated.
- The error between the set phase difference and the actual phase difference is measured and output to the RF main controller.
- Maintainability and modularity of the system was given importance so the staff can replace any damaged components with ease.
- Various other subsystems of the cyclotron are also developed in 80s and need to be upgraded.
- Future work consists of upgrading these systems with a more compact and robust design, similar to the work that was done in this project..



ACKNOWLEDGEMENTS AND REFERENCES

We want to give a special thanks to the UWMCF staff: Marissa Kranz, Rob Emery, and Eric Dorman. Thank you for your patience, expertise and support.

- [1] David Cuadrado Calle, José Antonio López Pérez, D. C. Calle, and J. A. López Pérez, "Gain And Phase Detector Based On The Analog Devices AD8302 Chip," INFORME TÉCNICO, Mar. 2012.
- [2] "Accurate Gain/Phase Measurement at Radio Frequencies up to 2.5 GHz | Analog Devices."
- [3] "AD8302 Datasheet and Product Info | Analog Devices."
- [4] P. Hiscoks, "Analog Circuit Design,"
- [5] B. Razavi, Design of analog CMOS integrated circuits, Second edition. New York, NY: McGraw-Hill Education, 2017.